

TEST REPORT



Report No. : 17-034380-01-2

Page of Pages : (1) / (5)



1. Client

Name : YOUNG TECH Co., Ltd

Address : 81, Hwanggeum-ro 89beon-gil, Hagun-ri, Yangchon-eup, Gimpo-si, Gyeonggi-do, Korea

Date of Receipt : 2017. 06. 05

2. Use of Report : For Quality Control

3. Test Sample

Description : IP CONVERTER

Manufacturer : YOUNG TECH Co., Ltd.

Model Name : YT-930

Serial Number : -

Remark : -

4. Date of Test : 2017. 06. 15. ~ 2017. 06. 16.

5. Test Standard/Method : IEC 60529: 2001

6. Testing Environment : Temperature : (24.1 ± 2.0) °C , Humidity : (37 ± 2) % R.H.

7. Test Results : Pass (IP66)

Note : 1. The test results contained apply only to the test sample(s) supplied by the client
2. This test report shall not be reproduced in full or in part without approval of the KTL in advance.

Affirmation	Tested by	Technical Manager
	Name : Chae Hui-dong (Signature)	Name : Kim Ji-hoon (Signature)

The above test report is the accredited test result by Korea Laboratory Accreditation Scheme, which signed the ILAC-MRA

2017. 06. 21

Korea Testing Laboratory

Accredited by KOLAS, Republic of KOREA




87, Digital-ro 26-gil, Guro-gu, Seoul, KOREA Tel.+82-2-860-1537 Fax. +82-2-860-1549

FP202-03-03

* 위 마크는 추후 전자확인증 대조 프로그램에서 원본대조시 사용되는 2D코드입니다.

<Contents>

1. Summary of Test	3
1.1 Test Standard	3
1.2 Test Sample	3
1.3 Test Environment	3
1.4 Remark	3
2. Results	4
2.1 Dust Test Conditions	4
2.2 Dust Test Contents	4
2.3 Water Test Conditions	4
2.4 Water Test Contents	4
3. List of Testing Equipments	4
4. Test Figures	5



1. Summary of Test

1.1 Test Standard

This test was conducted in accordance with "IEC 60529: 2001".

1.2 Test Sample

- Description : IP CONVERTER
- Model Name : YT-930
- Dimensions : 131 mm × 124.2 mm × 123.7 mm



[Fig. 1: Front]



[Fig. 2: Side]

1.3 Test Environment

- Temperature : (24.1 ± 2.0) °C
- Humidity : (37 ± 2) % R.H.
- Atmospheric Pressure : (100.1 ± 2.0) kPa

1.4 Remark

N/A.

2. Results

Code Letters	IP	Conditions	Results
1st Characteristic numerals Against ingress of solid foreign objects	6	2.1 Dust Test Conditions <ul style="list-style-type: none"> Talcum powder(mesh) wire diameter: 50 μm Talcum powder(mesh) wire width: 75 μm Amount of talcum powder of the test chamber: 2 kg/m^3 2.2 Dust Test Contents <ul style="list-style-type: none"> Volume of the enclosures: About 700 cm^3 Reduction air pressure: -2.00 kPa (-200 mmH_2O) Flow rate: 0.00 L/min Extraction rate per hour: 0 volumes/h Test duration: 8 h 	Pass
2nd Characteristic numerals Against ingress of water with harmful effects	6	2.3 Water Test Conditions <ul style="list-style-type: none"> Internal diameter of the nozzle: 12.5 mm Delivery rate: (100 \pm 5) L/min Core of the substantial stream: Circle of 120 mm diameter at 2.5 m distance from the nozzle Distance from nozzle to enclosure surface: 2.8 m 2.4 Water Test Contents <ul style="list-style-type: none"> Test duration: 3 min 	Pass

3. List of Testing Equipments

Equipment	Manufacture	Model	ICP No.	Date of Calibration	Calibration Laboratory
Thermo-hygrometer	TESTO	Testo 622	ICP20140892	2017. 05. 23	KTL
Vernia Caliper	MITUTOYO	CD-20APX	ICP20160207	2017. 05. 18	KTL
Flow Meter (Nozzle)	KOMETER	GA-101	ICP20150325	2017. 05. 31	DDHT
Flow Meter (Compact Chamber)	DWYER	RMA-13-SSV	ICP20150336	2017. 05. 25	KTL

FP202-02-02

4. Test Figures



[Fig. 3: IP 6X]



[Fig. 4: IP X6]

복사본 COPY

- End -

COPY 복사본

